

FOURIER TRANSFORM INFRARED (FT-IR) SPECTROSCOPY OF ATMOSPHERICALLY SIGNIFICANT MOLECULES^a

C. HARIDASS, J. JORDAN, H. LAUZIÈRE, M. KAMAL, and P. MISRA, *Laser Spectroscopy Laboratory, Department of Physics and Astronomy, Howard University, Washington D.C. 20059.*

Vibrational and rotational spectra of atmospherically significant molecules, like NO₂, SO₂, and HCl, in gas phase, have been recorded when the gases are passed through tubes made of different materials (e.g. stainless steel, aluminum, copper and teflon) under quasi-static and flow conditions, using a Nicolet Magna-IR 500 Fourier Transform Infrared spectrometer. Fundamental modes of vibrations of the triatomic molecules, NO₂ and SO₂, in the spectral region 400 - 4000 cm⁻¹ have been identified for absorption monitoring. The infrared spectra of some molecules, e.g. NO₂ and HCl, show measurable difference in absorbance peaks, when the gases are passed through copper tubing for instance, confirming that these molecules react strongly with metallic materials, whereas teflon and stainless steel are fairly inert. Our objective is to identify and characterize the various tubing materials with respect to their absorption features, so that specific tubing-gas combinations may be determined for designing suitable platforms for airborne atmospheric measurements.

^aFinancial support from the U.S. Environmental Protection Agency's Office of Exploratory Research (#R819720-02-0) and the Center for the Study of Terrestrial and Extraterrestrial Atmospheres (NASA NAGW-2950) is gratefully acknowledged.