

COAXIAL PULSE JET SUB-MILLIMETER WAVE FAST SCAN SPECTROMETER WITH A FREQUENCY AND PHASE STABILIZED BWO

JOHN W. BEVAN, ROBERT R. LUCCHESE AND BLAKE A. MCELMURRY, *Department of Chemistry, Texas A&M University, College Station, TX 77843-3255*; SERGEY P. BELOV and IGOR I. LEONOV, *Institute of Applied Physics of Russia Academy of Sciences, Uljanova Str. 46, Nizhny Novgorod 603600, Russia*.

A coaxially configured pulsed jet sub-mm wave spectrometer with a digitally controlled fast scan frequency and phase stabilized Backward Wave Oscillator (BWO) has been constructed at TAMU. The spectrometer samples a segment of high resolution absorption spectra of jet-cooled species during a single jet pulse (~ 1 ms) by digitally scanning the BWO frequency with $10 \mu\text{s}$ (current minimum time) per step. The value of frequency step can be varied from 10 Hz to 100 kHz. To reduce a systematical error in frequency measurements, BWO can be scanned up and down during a given nozzle pulse. Newly developed software allows continuously coverage of the whole frequency range of the BWO using coadded frequency segments with signal averaging of a number (~ 1000) of pulses for each segment of recorded spectrum. Test measurements of jet-cooled spectra of monomers (CO, DBr) and molecular complexes (Ar-HBr) has shown that the accuracy of frequency measurements and resolution of this direct absorption spectrometer both are very similar (in a relative units) to those of pulse-nozzle Fourier transform MW spectrometers with a parallel beam configuration. The actual frequency range of the spectrometer 240 - 380 GHz (BWO OB-30) can be extended to 1 THz by simple replacement of BWO. A block-diagram of the spectrometer, its parameters and some jet-cooled spectra will be presented.