

THE MEASUREMENT OF AMBIENT NO₂ BY CAVITY RING-DOWN SPECTROSCOPY

JAMES HARGROVE, LIMING WANG, AIKO MATSANUGA, SUSAN ZAIDE, and JINGSONG ZHANG,
Department of Chemistry and Air Pollution Research Center, University of California, Riverside, CA 92521;
KAREN MUYSKENS and MARK MUYSKENS, *Department of Chemistry, Calvin College, Grand Rapids,*
MI 49546.

Nitrogen dioxide (NO₂) is a principle component of photochemical smog that is present in polluted atmospheres at the parts per billion (ppb) level. In this work, cavity ring-down spectroscopy (CRDS) is utilized for the detection of ambient NO₂. Sub ppb limit of detection is routinely achieved. Experimental approaches to remove the particulates and to improve the detection limit and accuracy are investigated. Ambient measurements of NO₂ by CRDS are evaluated, and are compared side by side with the NO₂ measurements by chemiluminescence analysis.