

LOW UNCERTAINTY MEASUREMENTS OF LINE PARAMETERS USING FREQUENCY-STABILIZED CAVITY RING-DOWN SPECTROSCOPY

J. T. HODGES, *National Institute of Standards and Technology, Gaithersburg, MD 20899*; D. LISAK, *Nicolaus Copernicus University, Torun, Poland*; D. J. ROBICHAUD and M. OKUMURA, *Division of Chemistry and Chemical Engineering, California Institute of Technology, Pasadena CA 91125*; C. E. MILLER, *Jet Propulsion Laboratory, Pasadena, CA 91109*.

We discuss how the frequency-stabilized cavity ring-down spectroscopy (FS-CRDS) method enables high-precision and high-sensitivity measurements of shapes, intensities, as well as broadening and shifting parameters of absorption lines^a. We illustrate the method's high spectral resolution by observation of sub-MHz wide Doppler-free saturation features in blended near-ir H₂O spectra, and we present measurements and models of H₂O and O₂ spectra that yield line parameters with sub%-level relative uncertainties.

^aD. Lisak, J. T. Hodges and R. Ciurylo *Phys. Rev. A*. 73, 012507 2006.