

## PROGRESS TOWARDS CHIRPED-PULSE FOURIER TRANSFORM THZ SPECTROSCOPY

KEVIN O. DOUGLASS, DAVID F. PLUSQUELLIC, *NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY, OPTICAL TECHNOLOGY DIVISION, GAITHERSBURG, MD 20899*; EYAL GERECHT, *NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY, ELECTROMAGNETICS DIVISION, BOULDER, CO 80305*.

New opportunities are provided by the development of higher power THz frequency multiplier sources, the development of a broadband Chirped-Pulse FTMW spectroscopy technique at microwave<sup>a</sup> and mmWave frequencies, and recently demonstrated heterodyne hot electron bolometer detection technology in the THz frequency region with near quantum noise-limited performance and high spectral resolution.<sup>b</sup> Combining these three technologies and extending the chirped-pulse technique to 0.85 THz enables a host of new applications. NIST is currently pursuing applications as a point sensor for greenhouse gases, volatile organic compounds, and potentially human breath. The generation and detection of phase stable chirped pulses at 850 GHz will be demonstrated. A description of the experimental setup and preliminary data will be presented for nitrous oxide.

---

<sup>a</sup>G.G. Brown, B.C. Dian, K.O. Douglass, S.M. Geyer, S. Shipman and B.H. Pate, *Rev.Sci.Instrum.* 79 (2008) 053103.

<sup>b</sup>E. Gerecht, D. Gu, L. You, K.S. Yngvesson, *IEEE TRANSACTIONS ON MICROWAVE THEORY AND TECHNIQUES.* 56, (2008) 1083.