

OH DETECTION USING OFF-AXIS INTEGRATED CAVITY OUTPUT SPECTROSCOPY (OA-ICOS)

CHRISTOPHE LENGIGNON^a, WEIXIONG ZHAO^b, WEIDONG CHEN, ERIC FERTEIN, CECILE COEUR, *Laboratoire de Physico-Chimie de l'Atmosphere, Universite du littoral Cote d'Opale, Dunkerque - France*; DENIS PETITPREZ, *Laboratoire de Physicochimie des Processus de Combustion et de l'Atmosphere, Universite des Sciences et Technologies de Lille, 59655 Villeneuve d'Ascq Cedex - France*.

The OA-ICOS cavity consisted of two 1 high reflectivity spherical mirrors with 1 m radius of curvature, separated by a 0.5-m long quartz coated stainless steel tube. The mirrors reflectivity was $> 99.996\%$ at 1435 nm as specified by the manufacturer (Layertec, GmbH). The effective optical path length of the OA-ICOS approach was determined with direct absorption signal intensity of the pure H₂O vapor line at 6965.80233 cm⁻¹ and was found to be 1.263 km.

The OH radicals were generated with the help of a 2.45 GHz microwave discharge in water vapor flow under low pressure (1 mbar) to evaluate the developed OA-ICOS performance. The OH radical concentration of 7.28×10^{13} OH/cm³ was determined using calibration with a close H₂O absorption line at 6965.80 cm⁻¹. The detection limit, deduced from the signal to noise ratio, was 3.86×10^{11} OH/cm³. Experimental instrument detail and the preliminary measurement results will be presented and discussed.

^aThis work is supported by the IRENI program of the Region Nord-Pas de Calais. The support of the Groupement de Recherche International SAMIA between CNRS (France), RFBR (Russia) and CAS (China) is acknowledged.

^bthanks the IRENI program for the postdoctoral support.